

Airborne Fiber Analysis

NIOSH 7400 Method, Issue 2, 15 August 1994, counting rules 'A'

Sample Company
Sample Contact
000 Sample Street
Suite 000
Sample City, CA 00000

Client ID: 0000
Report Number: A000000
Date Received: 00/00/0000
Date Analyzed: 00/00/0000
Date Printed: 00/00/0000
First Reported: 00/00/0000

Job ID/Site: Sample Site

FALI Job ID: 0000
Total Samples Submitted: 5
Total Samples Analyzed: 5

Sample ID	Lab Number	Date Collected	Volume (L)	Fibers	Fields	Fibers/mm ²	LOD F/cc	Fibers/cc
001	00000000	00/00/00	996.5	0.0	100	<7.0	0.003	< 0.003
002	00000000	00/00/00	929.3	0.0	100	<7.0	0.003	< 0.003
003	00000000	00/00/00	931.5	0.0	100	<7.0	0.003	< 0.003
004	00000000	00/00/00	926.2	0.0	100	<7.0	0.003	< 0.003
005	00000000	00/00/00	0.0	0.0	100	--	--	--

Comments: This result was used to blank correct the other samples on this report. Blank filters are reported only as number of fibers and fields counted.

Laboratory Supervisor, Hayward Laboratory

Intralaboratory Relative Standard Deviation (Sr) per 100 graticule fields: 5 to 20 fibers: 0.470; >20 to 50 fibers: 0.387; >50 to 100 fibers: 0.405

Analytical results and reports are generated by Forensic Analytical Laboratories Inc. (FALI) at the request of and for the exclusive use of the person or entity (client) named on such report. Results, reports or copies of same will not be released by FALI to any third party without prior written request from client. This report applies only to the sample(s) tested and results are based upon sample information provided by the client. Supporting laboratory documentation is available upon request. This report must not be reproduced except in full, unless approved by FALI. The client is solely responsible for the use and interpretation of test results and reports requested from Forensic Analytical. This report must not be used by the client to claim product endorsement by NVLAP or any other agency of the U.S. Government. FALI is not able to assess the degree of hazard resulting from materials analyzed. FALI reserves the right to dispose of all samples after a period of thirty (30) days, according to all state and federal guidelines, unless otherwise specified. Samples are not blank corrected unless otherwise noted. All samples were received in acceptable condition unless otherwise noted.