



Metals Analysis of HUD Wipes

Sample Company
Sample Contact
000 Sample Street
Suite 000
Sample City, CA 00000

Client ID: 0000
Report Number: M000000
Date Received: 00/00/0000
Date Analyzed: 00/00/0000
Date Printed: 00/00/0000
First Reported: 00/00/0000

Job ID / Site: Sample Site
Date(s) Collected: 00/00/0000

SGSFL Job ID: 0000-0
Total Samples Submitted: 3
Total Samples Analyzed: 3

Sample Number	Lab Number	Area ft2	Analyte	Result	Result Units	Reporting Limit*	Method Reference
1	00000000	1.00	Pb	24	ug/ft2	8	NIOSH 9100/7082
2	00000000	1.00	Pb	< 8	ug/ft2	8	NIOSH 9100/7082
3	00000000	1.00	Pb	< 8	ug/ft2	8	NIOSH 9100/7082

Note to clients performing work related to the Lead Based Paint Hazard Reduction Act: Sample results for wipes not meeting ASTM E 1792 are not recognized within the National Lead Laboratory Accreditation Program.

Forensic Analytical can not determine whether or not wipes submitted to us for analysis meet the ASTM standard. We recommend to our clients that they document the brand of wipe that they use for each submission on their sample request form.

* The Reporting Limit represents the lowest amount of analyte that the laboratory can confidently detect in the sample, and is not a regulatory level. The Units for the Reporting Limit are the same as the Units for the Final Results.

Laboratory Supervisor, Hayward Laboratory

Analytical results and reports are generated by SGS Forensic Laboratories (SGSFL) at the request of and for the exclusive use of the person or entity (client) named on such report. Results, reports or copies of same will not be released by SGSFL to any third party without prior written request from client. This report applies only to the sample(s) tested. Supporting laboratory documentation is available upon request. This report must not be reproduced except in full, unless approved by SGSFL. The client is solely responsible for the use and interpretation of test results and reports requested from SGSFL. SGSFL is not able to assess the degree of hazard resulting from materials analyzed. SGSFL reserves the right to dispose of all samples after a period of thirty (30) days, according to all state and federal guidelines, unless otherwise specified. Any modifications that have been made to referenced test methods are documented in SGSFL's Standard Operating Procedures Manual. Sample results have not been blank corrected. Quality control and sample receipt condition were acceptable unless otherwise noted.